



## ***Idaho State Police Forensic Services***

### **CERTIFICATE OF ANALYSIS**

The Department of Law Enforcement Idaho State Police Forensic Services hereby certifies that Simulator Solution Lot Number 0000006104 to be used to conduct calibration checks within the State of Idaho in accordance with the policies and/or procedures promulgated by the Department governing breath alcohol examinations. This lot has a target value of 0.200 with a range of 0.180 to 0.220 grams of ethyl alcohol/210 liters of vapor.

The expiration date for this lot number is March 4, 2008 at 11:59 PM.

October 23, 2006  
Date

\_\_\_\_\_  
Forensic Scientist

STATE OF IDAHO     )  
                                  ) ss.  
County of Ada    )

On this 23<sup>rd</sup> of October, in the year 2006, before me, Judy Packer, a notary public, personally appeared David A. Laycock, known to me to be the person whose name is subscribed to the within instrument as a Forensic Scientist for the Idaho State Police Forensic Services, and acknowledge to me that he executed the same as such Scientist.

\_\_\_\_\_  
Notary Public

Commission Expires: \_\_\_\_\_